

MIROMESH[®] BIOLOGIC MATRIX

MIROMESH[®] BIOLOGIC MATRIX INSTRUCTIONS FOR USE

CAUTION - Rx ONLY

Federal (U.S.A.) law restricts this device to sale by or on the order of a physician.

This product is intended for use by trained medical professionals.

DEVICE DESCRIPTION

MIROMESH Biologic Matrix is a non-crosslinked acellular surgical mesh that is derived from the highly vascularized porcine liver and is processed and stored in a phosphate buffered aqueous solution. MIROMESH Biologic Matrix is supplied sterile, packaged in an inner sterile pouch and outer non-sterile pouch, and is intended for one-time use. In clinical use, the device should be placed in maximum possible contact with healthy, well-vascularized tissue to assist in initiating cell ingrowth and tissue remodeling.

MIROMESH is available in six sizes listed in Table 1. Each device is packaged individually.

Table 1: MIROMESH Size Offerings

Size (dimensions)	Model Number	Quantity
6 x 8 cm	BLM-100-01-0608	1
8 x 8 cm	BLM-100-01-0808	1
10 x 10 cm	BLM-100-01-1010	1
8 x 16 cm	BLM-100-01-0816	1
10 x 16 cm	BLM-100-01-1016	1
10 x 20 cm	BLM-100-01-1020	1

INDICATIONS FOR USE

The MIROMESH Biologic Matrix is intended to be implanted to reinforce soft tissue and is also intended for implantation to reinforce soft tissue where weakness exists in patients requiring soft tissue repair or reinforcement in plastic and reconstructive surgery.

CONTRAINDICATIONS

This device is derived from a porcine source and should not be used in patients with known sensitivity to porcine material.

WARNINGS

- MIROMESH Biologic Matrix is supplied sterile for single use only. Reuse of a single-use device creates a potential risk of patient user infections and may compromise the device functionality, which may lead to illness or serious injury.
- Do not resterilize as the safety and performance has not been evaluated for this scenario.
- Do not use a device past the expiration date as the safety and performance has not be evaluated.
- Do not use if the package or seal is opened, damaged, or compromised. A damaged package could result in a breach of sterility or device damage.
- Do not use the product if the heat indicator has been activated as the safety and performance of the device has not been evaluated for that scenario.
- After use, handle and dispose of all unused product and packaging in accordance with accepted medical practice and applicable local, state and federal laws and regulations. This is a single-use device. Reuse of this device creates a potential risk of patient infections.

PRECAUTIONS

- Discard device if mishandling has caused possible damage or contamination as it may have resulted in breach of sterility or compromised device functionality.
- Soak the device for a minimum of 2 minutes using a sterile basin and room temperature sterile saline or room temperature sterile lactated Ringer's solution to cover the mesh to assist in removing some of the phosphate buffered aqueous solution in which the device is packaged.
- MIROMESH Biologic Matrix should be hydrated and moist when the package is opened. If the device is dry, do not use as a dry device as it may compromise device functionality.

STORAGE

- MIROMESH Biologic Matrix is a sterile medical device that should be stored in a clean, dry location at room temperature, in its original package. Avoid prolonged exposure to elevated temperatures as it may compromise device functionality.

- The product expiration date is indicated as year (4 digits) and month (2 digits). The product expires after the last day of the month indicated.

STERILIZATION

This product has been sterilized with electron beam irradiation and is provided sterile.

MR SAFETY INFORMATION

The MIROMESH Biologic Matrix is MR safe.

INSTRUCTIONS FOR USE

The MIROMESH Biological Matrix should be used by physicians trained on the procedures for which the device is intended. The techniques and procedures described do not represent ALL medically acceptable protocols, nor are they intended as a substitute for the physician's experience and judgment in treating any specific patient. All available data, including the patient's signs and symptoms and other diagnostic test results, should be considered before determining a specific treatment plan.

PACKAGE CONTENTS: One (1)

MIROMESH Biologic Matrix device in a phosphate buffered aqueous solution

REQUIRED MATERIALS NOT PROVIDED

- Sterile forceps.
- Soaking solution: room temperature sterile saline or room temperature sterile lactated Ringer's solution.
- One sterile basin per piece of MIROMESH Biologic Matrix.

PREPARATION INSTRUCTIONS

1. Check the heat indicator on the carton. Do not use the product if the heat indicator is activated.
2. Open the carton and remove the foil pouch.
3. Inspect the outer foil pouch. Do not use if the outer pouch is damaged or if the seals are compromised.
4. Peel open the outer foil pouch and remove the inner foil pouch using aseptic technique. The inner foil pouch is sterile and may be placed directly into the sterile field.
5. Peel open the inner foil pouch and remove the product using aseptic technique. Always use sterile gloved hands or forceps when handling the product.
6. Soak the device for a minimum of 2 minutes using a sterile basin and room temperature sterile saline or room temperature sterile lactated Ringer's solution to cover the mesh.
7. Store the device in the room temperature sterile solution until ready for implantation. The device can be stored in sterile solution for a maximum of 4 hours.



2725 12th Street East
Glencoe, MN 55336

952-377-8238

WWW.REPRISEBIO.COM

PROLENE[®] is a registered trademark of
ETHICON, a Johnson & Johnson
Company.

SP-1036, Rev D
August 2019

IMPLANTATION INSTRUCTIONS

1. Prepare the graft site using standard surgical techniques.
2. Using aseptic technique, MIROMESH Biologic Matrix may be trimmed or cut as desired to fit the surgical site, ensuring allowance for overlap.
3. MIROMESH Biologic Matrix should be placed in maximum possible contact with healthy, well-vascularized tissue to assist in initiating cell ingrowth and remodeling.
4. Suture MIROMESH Biologic Matrix in place; when suturing place the sutures at least 3 mm from the edge of the MIROMESH Biologic Matrix.
5. Complete the standard surgical procedure per institutional protocol.
6. Discard any unused portion of the MIROMESH Biologic Matrix product and package in accordance with accepted medical practice and applicable local, state and federal laws and regulations.


INSTRUCTIONS FOR QUILTING (I.E., SUTURING TOGETHER) TWO OR MORE MIROMESH BIOLOGICS MATRICES

NOTE: If a larger size of MIROMESH Biologic Matrix is needed, two or more MIROMESH Biologic Matrices may be sutured together to a maximum size of 20 cm x 40 cm quilted mesh using the instructions below.

WARNING

The resulting quilted mesh line must be placed perpendicular to the defect as described below and shown in Figure 1 for three quilted meshes.

Figure 1: Orientation and Placement Diagram for Three Quilted Matrices



QUILTING PRECAUTIONS

- Quilting of two or more pieces of MIROMESH Biologic Matrix is not recommended for any bridging procedures.
- Do not allow the meshes to dry out while suturing them together.
- When trimming quilted meshes, do not cut or remove the suture knots.

ADDITIONAL REQUIRED MATERIALS FOR SUTURING TWO OR MORE MIROMESH BIOLOGIC MATRICES – NOT INCLUDED

- CT-1 surgical needle with non-absorbable suture (such as PROLENE®) Size 0, minimum length of 30"
- Sterile ruler


SUTURING INSTRUCTIONS:

1. Using aseptic technique, align and stack two MIROMESH Biologic Matrices along edges of equal length per Figure 2. Place a sterile ruler along the matrices 1 cm from the edge of the top MIROMESH Biologic Matrix as a suture guide. All running sutures require a 1 cm bite (Figure 2).

NOTE: It may be helpful to apply gentle tension on opposite ends of the stacked meshes during quilting to keep the meshes from moving.

2. Anchor a suture on one end of the stacked MIROMESH Biologic Matrices with a minimum of six surgeon's knots and suture the two pieces of mesh together with a running suture every 1 cm (Figure 2). Suture line should be pulled taught during suturing process to ensure uniform tension along the line. Once the edge of the stacked meshes has been reached, secure the suture with a minimum of six surgeon's knots. Cut the suture to leave a tail of appropriate length. **NOTE:** The surgeon may add additional throws to each knot as necessary.

Figure 2: Diagram of Two Stacked Matrices with the First Running Suture




3. Beginning at the opposite end of the stacked MIROMESH Biologic Matrices, anchor another suture with a minimum of six surgeon's knots and **suture the two pieces of mesh together with a running suture in the opposite direction every 1 cm in between the spaces formed by the first running suture resulting in a final suture spacing of 0.5 cm (Figure 3).**

Secure with a minimum of six surgeon's knots at the opposite end. Cut the suture to leave a tail of appropriate length.

NOTE: The surgeon may add additional throws to each knot as necessary.

Figure 3: Diagram of Two Stacked Matrices with Both Running Sutures



4. The resulting quilted line will have a crisscross appearance to the sutures and a 0.5 cm spacing between the sutures (Figure 4 & 5).

Figure 4: Diagram of Finished Matrices





Figure 5: Side View of Example Quilted Matrix



5. Open the quilted mesh so that it is a single layer and observe the crisscross pattern (see Figures 6 & 7). Position the mesh with the suture ridge anterior (visible to the surgeon) with the suture ridge running perpendicular to the incision upon implant (see Figure 8). Complete the standard surgical procedure.

Figure 6: Top View Example of Crisscross Pattern in Quilted Matrix




Figure 7: Top View Diagram of Two Quilted Matrices





Figure 8: Placement Diagram for Two Quilted Matrices



6. **When quilting together three or more matrices (maximum size is 20 cm x 40 cm), ensure each mesh is quilted such that the quilted mesh line will be oriented perpendicular to the defect line (see Figure 9).**
7. Once quilting is complete, open the quilted mesh so that it is a single layer. Position the mesh with the suture ridges anterior (visible to the surgeon) with the suture ridge running perpendicular to the incision upon implant (see Figure 9).
8. Complete the standard surgical procedure.

Figure 9: Placement Diagram for Three Quilted Matrices

